



BID BULLETIN NO. 1

**PROCUREMENT OF VARIOUS DEVICES AND EQUIPMENT FOR
PHYSICAL REHABILITATION, ENT AND ER USE**

REFERENCE: IB NO. 2022-06-07 (14)

***** AMENDMENT TECHNICAL SPECIFICATION *****

ORIGINAL DESCRIPTION	AMENDED DESCRIPTION
<p>THERAPEUTIC ULTRASOUND</p> <ul style="list-style-type: none">- 7 inches colored touchscreen display- Programmable, selectable coupling sensitivity- Dimension: 380x190x260 mm- Weight: 3 kg max- Power supply: 100-200 v/ac, 50-60 Hz- Modular Frequency: 10-150 Hz- Duty Factor: 5-100%- Max intensity continuous mode: 2 W/cm²- Max intensity pulse mode: 3 W/cm²	<p>1 Channel Ultrasound - Ultrasound Head: at least 5 cm and with free 1 cm ultrasound head</p> <ul style="list-style-type: none">- at least 7 inches display screen with quick protocols-Programmable, selectable compiling sensitivity- Dimension: 380 x 190 x 260 mm-Weight: 3 kg max-With trolley-Main Supply: 100-240 v/ac, 50-60Hz-Modular Frquency: 10-150 Hz-Duty Factor: 5-100%-Max Intensity Continuous mode: 2 W/-Max Intensity Pulse Mode: 3 W/cm²

Please be guided accordingly.

HILARION P. PASAL, CE, SE
BAC Chairman