



## BID BULLETIN NO. 1

### PROCUREMENT OF HEMODIALYSIS SUPPLIES


IB NO. 2022-06-28 (19)

### **\*\*\* AMENDMENT OF TECHNICAL SPECIFICATION AND OPENING OF BIDS SCHEDULE \*\*\***

ITEM NO	ORIGINAL	AMENDED
1	HEMODIALYSIS BLOOD TUBING SET ACCESSORIES: PILLOW, TRANSDUCER PROTECTOR, IV SET NON-DEHP, NON-PYROGENIC, NON-TOXIC, KINK RESISTANT TUBING	HEMODIALYSIS BLOOD TUBING SET ACCESSORIES: PILLOW, TRANSDUCER PROTECTOR, <i>DEHP OR NON-DEHP OR MEDICAL GRADE</i> , NON-PYROGENIC, NON-TOXIC, KINK RESISTANT TUBING
2	BICARBONATE CARTRIDGE	BICARBONATE <i>PREMIXED</i> <i>UNIT OF MEASURE : GALLOON</i>
3	ACID SOLUTION 5 LITERS (GALLON) CONCENTRATED HEMODIALYSIS SOLUTION	ACID SOLUTION ( <i>2 GALLONS:5 LITERS</i> ) CONCENTRATED HEMODIALYSIS SOLUTION
4	DIALYZER HIGH FLUX WITH A SURFACE AREA OR 1.9 SQ.M. AND WALL THICKNESS/OR INTERNAL DIAMETER OF 30 TO 40/200 MICROMETER	DIALYZER HIGH FLUX WITH A SURFACE AREA OR <i>1.8 SQ.M.</i> AND WALL THICKNESS/OR INTERNAL DIAMETER OF 30 TO 40/200 MICROMETER
5	DIALYZER LOW FLUX W/ A SURFACE AREA OF 1.9 SQ.M. AND WALL THICKNESS/INTERNAL DIAMETER OF 30-40/200 MICRO METER	DIALYZER LOW FLUX W/ A SURFACE AREA OF <i>1.8 SQ.M.</i> AND WALL THICKNESS/INTERNAL DIAMETER OF 30-40/200 MICRO METER

**Opening of Bids Schedule: JUNE 30, 2022, 9:00 A.M.**

Please be guided accordingly.

  
**HILARION P. PASAL, CE, SE**  
BAC Chairman