



SUPPLEMENTAL BID BULLETIN NO. 1


**PROCUREMENT OF INFUSION PUMPS FOR PEDIA DEPARTMENT
IB NO. 2023-04-24 (15)**

***** AMENDED SPECIFICATION *****

The following details are the final specification set by this Procuring Entity as amended during the Pe-Bidding Conference to wit:

ORIGINAL SPECIFICATION POSTED	AMENDED SPECIFICATION
Should have a minimum 5-7 hours battery back up at highest delivery rate.	Should have a minimum of <u>4.5 – 7 hours and above</u> battery back up at highest delivery rate
AC Power Supply: 220-230 V / 50-60 Hz. With option for external DC power supply (13-15 Volts) when used to transport patient via ambulance	AC Power Supply: 220-230 V / 50-60 Hz. With option for external DC power supply (<u>12-15 Volts</u>) when used to transport patient via ambulance

Please be guided accordingly.


ATTY. MA. FLORDELIN T. ENSOMO, CPA
Attorney IV
Chairman, Bids and Awards Committee