



SUPPLEMENTAL BID BULLETIN NO. 1


PROCUREMENT OF ONE (1) UNIT ANESTHESIA MACHINE FOR OBGYNE DEPARTMENT IB NO. 2023-05-22 (25)

***** AMENDMENT OF TECHNICAL SPECIFICATION *****

The following details are the final specification set by this Procuring Entity as amended during the Pe-Bidding Conference to wit:

ORIGINAL SPECIFICATION POSTED	AMENDED SPECIFICATION
<ul style="list-style-type: none">• Must be electronically driven piston ventilator (E-Vent plus), fresh-gas decoupled, ventilation without drive gas, i.e. no medical gases are consumed in operating the ventilator (regardless of gas supply).	<ul style="list-style-type: none">• Must be electronically driven piston ventilator (E-Vent plus) Or Fully integrated, Electronically controlled, pneumatically driven OR its equivalent, fresh-gas decoupled, ventilation without drive gas, i.e. no medical gases are consumed in operating the ventilator (regardless of gas supply).
<ul style="list-style-type: none">• The monitoring solution's battery shall support a minimum of 350 minutes of continuous monitoring	<ul style="list-style-type: none">• The monitoring solution's battery shall support a minimum of 150 minutes of continuous monitoring
A. Medline Adult Expandable Anesthesia Circuits Reusable, 20 Each per Case	A. Adult Expandable Anesthesia Circuits Disposable , 20 Each per Case
B. Medline Pediatric Expandable Anesthesia Circuits Reusable, 20 Each per Case	B. Pediatric Expandable Anesthesia Circuits Disposable , 20 Each per Case
Addition:	The Supplier shall include sevoflurane vaporizer.

Please be guided accordingly.


ATTY. MA. FLORDELIN T. ENSOMO, CPA
Attorney IV
Chairman, Bids and Awards Committee